



Options	CTC256	TTC256	Order no.
Works calibration certificate for one temperature according to customer specification (only TTC)	-		D00109
Works calibration certificate for one temperature and humidity according to customer specification (only CTC)		-	D00105
Door-hinges on left side			B8
Full-sight glass door (5-layer insulating glazing), heated			B0
Entry port, left, 80 mm, with stopper			F0
Interface Ethernet instead of USB including software			W4
RS232 interface instead of USB			W6
ditto, for combination error message (e.g. supply failure, sensor fault, fuse); VO only in combination with option T5			H6
MobileALERT, notification by SMS in case of any error or alarm of the device. Requires option H6 "floating contact for alarm"			C3

Accessory	CTC256	TTC256	Order no.
Additional stainless steel grid			E20591
External control and logging package consisting of mini-Notebook and software „Celsius“, pre-configured, and lateral swinging arm			B04410
USB connection cable for computer interface			E03643
Temperature profile write/read unit for programming via PC, for writing to and reading from the chip card, up to 40 ramps			E05284
Additional chip card, blank, formatted (32 kB MEMoryCard XL for a maximum of 40 ramps)			E04004
Oven-linked authorisation card (User-ID-Card) prevents undesired manipulation by unauthorised third parties. When reordering please specify serial number			E04159
Software conforming to FDA "Celsius FDA Edition" for up to 16 units. Meets the requirements for the use of electronically stored data sets and electronic signatures as laid down in Regulation 21 CFR Part 11 of the US Food and Drug Administration (FDA). Not applicable for UFP800DW			E05019
Integration of additional units (up to max. 16 units) into an already existent FDA-software licence. Not applicable for UFP800DW			FDAQ4
IQ check list with works test data for chamber as support for validation by customer			D00103
OQ check list with works test data for one free-selectable temperature value incl. temperature distribution survey for 27 measuring points to DIN 12880:2007-05 as support for validation by customer		-	D00144
External measuring instrument with sensors for daylight and UV-light, with additional measuring head for temperature and humidity. Product information on demand (models HCP and CTC)		-	B04714

